Fluke 77 Series III Front Panel Replacement

Front panel replacement for the Fluke 77 Series III Multimeter.

Written By: Kishan Patel



INTRODUCTION

This is a guide for replacing the front panel of the Fluke 77 Series III Multimeter. Front panel removal is a prerequisite-only guide for replacing all other parts of the Fluke 77 Series III Multimeter.

TOOLS:

- Phillips #2 Screwdriver (1)
- iFixit Opening Tools (1)

Step 1 — Front Panel



- ⚠ Be sure any test leads are out and the multimeter is off before opening the device.
- Turn the multimeter so it is display side down.
- Lift up the back stand of the device by putting your thumb in the curved well at the bottom of the device, so you have access to the two covered screws below.

Step 2



- Use your Phillips #2 Screwdriver to remove the four black 21 mm Phillips #2 screws.

Step 3





- Turn the Fluke 77 Series III Multimeter back over so that it is display side up.
- Pry the front panel from the back panel at the seam.
- If the front panel does not come off easily, move a plastic opening tool along the seam until the panels separate completely.

To reassemble your device, follow these instructions in reverse order.

This document was last generated on 2017-06-30 11:13:22 AM.